

For Immediate Release: July 29, 2014

INTERTEK WARNS OF NON-COMPLIANT ETL_{US} CERTIFICATION OF PORTABLE GAS DETECTORS FOR USE IN CLASS I, DIVISIONS 1 & 2, GROUPS A - D HAZARDOUS LOCATIONS

Intertek Testing Services (ETL) notifies consumers, retailers and distributors that the Portable Gas Detectors identified below may not be in compliance with their Listing Standards.

UL Standard for Safety for Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations [UL 913, Fifth Edition; Dated February 21, 1997; with Revisions through and including February 24, 1997]

UL Standard for Safety for Explosion Proof and Dust-Ignition-Proof Equipment for Use in Class I, Division 1 and Class I Division 1 Hazardous (Classified) Locations, UL 1203, 4th Edition with revisions October 28, 2009

UL Standard for Safety for Electrical Equipment for Measurement, Control and Laboratory Use - Part 1: General Requirements [ANSI/UL 61010-1-2008; Second Edition; Dated July 12, 2004]

The number of Portable Gas Detectors bearing the ETL_{US} Certification Mark that have been produced or released into the market is believed to be less than 250.

- Name of Product: Portable Gas Detectors**
- Model: DataFID LD, SiteFID, MicroFID II, & BT antenna equipped DATAFID LD, SiteFID, MicroFID II**
- Brand/Manufacturer: Inficon, Inc.**
- Manufacture Years: 2011-2014**
- ETL Control Number: 3079292**
- Sold: Inficon Direct sales to commercial/industrial customers**
- Claim: The integral rechargeable batteries have not been evaluated for use in the products**
- Recall Instruction: Contact Inficon**

Photos of Product:

